

Processing Record for this file:
(Make notes here unique for this file)

Loading from 2014-09-01 - 2014-10-01

Data taken from DMF, untarred, unzipped and renamed back to original file names


It was noted that BOTH TEI 42i's lost their internal Mass Flow Meters. This was not consequential as the inlet has much better Mass Flow Controllers.

Ambient data appear valid.


2014-09-14 a 47-mm x 5-um Teflon filter was added upstream of the NO/NO_x channel. There has been issues with the operation of the winch to raise/lower the NO_x inlet tower (necessary to change the filter on a 2-week schedule). Some evidence of reduced sensitivity after ~2014-09-24. It is not clear if this is due to the filter becoming dirty. The photolysis conversion efficiency is quite high (0.74) which is good. iAlso on 2014-09-24, the zero air signals on all three channel increased substantially.

Diurnal cycles of all three species are readily apparent. Generator operation and restarts seemed quite commone this month. No discrimination of these periods was provided.

Ground-Based NO_x Analyzer
Housekeeping 2
Temperatures



Ground-Based NO_x Analyzer
Housekeeping 1
Flows & Pressures




(UTC)

Project/Platform/Site: maomaoscs1
ARM Climate Research Facility
Contact: S.R. Springston (631) 344-4477, srs@bnl.gov

○ - NO Channel ; △ - NO_x Channel; □ - NO_y Channel
 Zero Air Equivalent Signal (ppbv)
 S.A. #1 (calibrated sig/dilution conc: [SA#1] = 21.51 ppbv)
 S.A. #2 (calibrated sig/dilution conc: [SA#2] = 85.01 ppbv)
 GPT + SA #2 (NO: frac titr., NO_x: Photol. C.E. (nom=0.7382), NO_y: Mo C.E. (nom=1.0085))
 HNO₃ cal. (calibrated sig/perm conc: [HNO₃] = 73.9 ppbv)


Ground NO_x Analyzer Calibration Data




Project/Platform/Site: maomaoscs1
 ARM Climate Research Facility
 Contact: S.R. Springston (631) 344-4477, srs@bnl.gov

(UTC)


Ground NO_x Analyzer Parsed NO_y Channel Data




Ground NO_x Analyzer Parsed NOx Channel Data



Ground NO_x Analyzer Parsed NO Channel Data




Calibration Factors Used:
NO: 1.07 ([calibrated]/[inst. reported])
NO₂: 1.087 ([calibrated]/[inst. reported]), C.E. = 0.7382
NO_y: 1.506 ([calibrated]/[inst. reported]), C.E. = 1.0085




Project/Platform/Site: maomaoscs1
ARM Climate Research Facility
Contact: S.R. Springston (631) 344-4477, srs@bnl.gov

Calibration Factors Used:
NO: 1.07 ([calibrated]/[inst. reported])
NO₂: 1.087 ([calibrated]/[inst. reported]), C.E. = 0.7382
NO_y: 1.506 ([calibrated]/[inst. reported]), C.E. = 1.0085




Project/Platform/Site: maomaoscs1
ARM Climate Research Facility
Contact: S.R. Springston (631) 344-4477, srs@bnl.gov


Calibration Factors Used:
NO: 1.07 ([calibrated]/[inst. reported])
NO₂: 1.087 ([calibrated]/[inst. reported]), C.E. = 0.7382
NO_y: 1.506 ([calibrated]/[inst. reported]), C.E. = 1.0085



Calibration Factors Used:
NO: 1.07 ([calibrated]/[inst. reported])
NO₂: 1.087 ([calibrated]/[inst. reported]), C.E. = 0.7382
NO_y: 1.506 ([calibrated]/[inst. reported]), C.E. = 1.0085



Calibration Factors Used:
NO: 1.07 ([calibrated]/[inst. reported])
NO₂: 1.087 ([calibrated]/[inst. reported]), C.E. = 0.7382
NO_y: 1.506 ([calibrated]/[inst. reported]), C.E. = 1.0085



Calibration Factors Used:
NO: 1.07 ([calibrated]/[inst. reported])
NO₂: 1.087 ([calibrated]/[inst. reported]), C.E. = 0.7382
NO_y: 1.506 ([calibrated]/[inst. reported]), C.E. = 1.0085

