

README

Passive Cavity Aerosol Spectrometer (PCASP)
GoAmazon IOP1 - February 22 to March 23, 2014
GoAmazon IOP2 - September 06 to October 04, 2014

Created by Jason Tomlinson for ARM Aerial Facility.
March 27, 2015

Contact:

Jason Tomlinson - jason.tomlinson@pnnl.gov

1 Data source

The data were recorded onboard the G1 aircraft, operated by the ARM Aerial Facility during GoAmazon IOP1 and IOP2

1.1 Location

The research flights were conducted out of the Eduardo Gomes-Manaus International Airport in Manaus, Brazil.


1.2 Time period of collection

Flights were conducted from February 22 - March 23, 2014 and September 06 to October 04, 2014.

1.3 Instrument description

Laser based instrument that uses the measured intensity of the scatter light from particles passing through the laser beam to measure aerosol concentration and size in the 0.1 to 3.45 μm size range. Instrument was calibrated during the field campaign using PSL (1.58 RI). The PCASP will report a large amount of false counts (at $D_p > 800\text{nm}$) in clouds or precipitation and end users should not use these data. A cloud flag has been included in the R1 data.

2 Data structure



2.1 Level R0

Level R0 data consist of raw data (counts) files and housekeeping files recorded at 1HZ.

2.2 Level R1

Level R1 data consist of a file containing integrated number, area, and volume concentrations. The mass flow controller has been corrected to volumetric flow using ambient pressure and temperature. Contains number distributions not normalized by dLogDp. All data is reported at 1HZ and contains a QC flag and cloud flag. File name includes takeoff time for the flight. Metadata has been added to the header of each file and follows the ICARTT standard.

<http://www-air.larc.nasa.gov/missions/etc/IcarttDataFormat.htm>

3 Data log

Date	Instrument			Notes
	Instrument Status	Data Process	QC	
20140222a	Ok	R1	Ok	
20140225a	Ok	R1	Ok	
20140301a	Ok	R1	Ok	
20140301b	Ok	R1	Caution	Instrument started about 17 minutes into flight
20140303a	Ok	R1	Caution	Instrument started about 25 minutes into flight
20140307a	Ok	R1	Ok	
20140310a	Ok	R1	Ok	
20140311a	Ok	R1	Ok	
20140312a	Ok	R1	Ok	
20140313a	Ok	R1	Ok	
20140314a	Ok	R1	Ok	
20140316a	Ok	R1	Ok	
20140317a	Ok	R1	Ok	
20140319a	Ok	R1	Caution	Data missing briefly around 16:40 UTC
20140321a	Ok	R1	Caution	Data missing from 18:15 to 18:35 UTC
20140323a	Ok	R1	Ok	
20140906a	Ok	R1	Ok	
20140909a	Ok	R1	Ok	
20140911a	Ok	R1	Ok	
20140912a	Ok	R1	Ok	
20140913a	Ok	R1	Ok	
20140915a	Ok	R1	Ok	
20140916a	Ok	R1	Ok	
20140918a	Ok	R1	Ok	
20140919a	Ok	R1	Ok	
20140921a	Ok	R1	Ok	
20140922a	Ok	R1	Ok	
20140923a	Ok	R1	Ok	
20140925a	Ok	R1	Ok	
20140927a	Ok	R1	Ok	
20140928a	BAD	R1	BAD	Loose wire in canister caused data to drop out for most of the flight
20140930a	BAD	R1	BAD	Loose wire in canister caused data to drop out for most of the flight
20141001a	Ok	R1	Ok	
20141003a	Ok	R1	Ok	
20141004a	Ok	R1	Ok	

4 File Format

File naming convention: "pcasp_g1_YYYYMMDDHHMMSS_Rx_Ly_goamazon001s.ict"

Where x is the revision number and y is the flight (launch) number for the day. The file is comma delimited. YYYYMMDD is the date of the flight and HHMMSS represents the takeoff time.

4.1 Data description

Index	Variable Name	Units	Range or	From Instrument:	Description Definition
			Frequency		
1	Start Time	UTC	1 s	PCASP	Seconds since midnight Synchronized with M300
2 to 31	Num_Conc	(#/cm ³)	1s	PCASP	Number Concentration at specified mean diameters
32	Num_Conc	(#/cm ³)	1s	PCASP	Integrated Number Concentration
33	Area_Conc	(mm ² /cm ³)	1s	PCASP	Integrated Area Concentration
34	Vol_Conc	(mm ³ /cm ³)	1s	PCASP	Integrated Volume Concentration
35	Data_Flag		0 to 2	PCASP	0: Good 1:Caution 2: Bad
36	Cloud Flag		0 to 2	IWG1	0: Good 1:Clouds may be present 2:Definite clouds

4.2 Definition and diagrams

QC	Description
Data_Flag	0: Good - Laser voltage > 7 and Num_Conc < CPC_Conc 1: Caution - 6 < Laser voltage < 7 or CPC_Conc < Num_Conc < CPC_Conc*1.25 2: Bad - Laser voltage < 6 or Num_Conc > CPC_Conc*1.25

4.3 Bins (μm)

Bin #	Lower	Upper	Mid	Bin #	Lower	Upper	Mid
1	0.090	0.100	0.095	16	0.300	0.400	0.350
2	0.100	0.110	0.105	17	0.400	0.575	0.488
3	0.110	0.120	0.115	18	0.575	0.690	0.633
4	0.120	0.130	0.125	19	0.690	0.920	0.805
5	0.130	0.140	0.135	20	0.920	1.150	1.035
6	0.140	0.150	0.145	21	1.150	1.380	1.265
7	0.150	0.160	0.155	22	1.380	1.610	1.495
8	0.160	0.170	0.165	23	1.610	1.840	1.725
9	0.170	0.180	0.175	24	1.840	2.070	1.955
10	0.180	0.200	0.190	25	2.070	2.300	2.185
11	0.200	0.220	0.210	26	2.300	2.530	2.415
12	0.220	0.240	0.230	27	2.530	2.760	2.645
13	0.240	0.260	0.250	28	2.760	2.990	2.875
14	0.260	0.280	0.270	29	2.990	3.220	3.105
15	0.280	0.300	0.290	30	3.220	3.450	3.335

5 Calibration

5.1 PSL

PSL calibration showed a consistent under sizing of the PSL by the PCASP in the low gain stage ($D_p > 0.450 \mu\text{m}$). For that reason, the bins were shifted by a factor of 1.15 starting with Bin #17.

